

## HeartSine® Mobile AED Rescue Backpack

For samaritan® PAD Automated External Defibrillators



\* Backpack sold separately. HeartSine offers the AED, Pad-Pak, Pediatric-Pak and CPR Prep Kit.

### Be Prepared for Cardiac Emergencies Even When You're on the Move

The HeartSine Mobile AED Rescue Backpack, custom made for the HeartSine samaritan Public Access Defibrillator (PAD), lets you create and transport a heart safe solution to virtually any location.

Sudden cardiac arrest (SCA) can't always be prevented, but immediate access to an AED can increase the chance of survival by up to 74%<sup>1</sup>, especially in remote areas where access to emergency services is often limited. In the event of an SCA, there's little time to react and even less time to think as an SCA can happen to anyone, at any time, regardless of age, gender or fitness.

The HeartSine Mobile AED Rescue Backpack has room for everything you should need to keep members of your group heart safe. This customized backpack can readily carry a HeartSine samaritan PAD AED with Pad-Pak™, a Pediatric-Pak™ or extra Pad-Pak, CPR prep kit and first aid kit.

#### Ideal for Active Organizations

- Mountain/Water/Disaster Rescue Teams
- Community First Responders
- Scouting/Youth Groups
- Cycling, Hiking and Running Clubs
- School Athletics and Outings
- Family Outings

And more...



#### Backpack Features

- Custom compartment for HeartSine AED with internal pocket for storage of Pad-Pak and Pediatric-Pak
- AED symbol for easy location of AED compartment
- First aid kit compartment with external symbol and small storage pocket
- D-rings for external attachment
- Large mesh pocket with adjustable cordlock to tighten
- Mesh side pockets with adjustable cordlock for storage of water bottle, cell phone and snacks
- Mesh shoulder strap pockets for flashlight, whistle or other small objects
- Adjustable shockcord with cordlock for keeping jacket within reach
- Adjustable chest and waist straps to fit adults or children
- Adjustable side straps to secure contents
- Interior mesh pocket to stabilize additional contents
- Padded back for extra comfort
- Adjustable padded shoulder straps
- Carrying handle
- Reflective safety strips

Brief summary of indications and important safety information on back.

### Why HeartSine AEDs?

- ▶ **Significantly smaller size** so it can travel with the group
- ▶ **Lightest AED** makes for easy transport in a backpack
- ▶ **Highest level of protection from dust and water (IP56)** offers unmatched robustness in both indoor and outdoor environments
- ▶ **Easy to use by all rescuers** due to simple one or two-button operation, intuitive voice/visual prompts and no lid to open nor complex displays or controls

### Backpack Specifications

- ▶ **Material:**  
Breathable polyester
- ▶ **Dimensions:**  
8.66 in w x 14.17 in h x 6.30 in d  
(220 mm w x 360 mm h x 160 mm d)
- ▶ **Weight:**  
Backpack only  
1.9 lbs (0.96 kg)  
  
Backpack with AED  
4 lbs (2.06 kg)
- ▶ **Color:**  
Red with black trim



# HeartSine® samaritan® PAD Automated External Defibrillators (AEDs)

## BRIEF SUMMARY OF INDICATIONS AND IMPORTANT SAFETY INFORMATION

### INDICATIONS FOR USE

The HeartSine samaritan PAD SAM 350P (SAM 350P), HeartSine samaritan PAD SAM 360P (SAM 360P) and HeartSine samaritan PAD SAM 450P (SAM 450P) are indicated for use on victims of cardiac arrest who are exhibiting the following signs: unconscious, not breathing, without circulation (without a pulse). The devices are intended for use by personnel who have been trained in their operation. Users should have received training in basic life support/AED, advanced life support or a physician-authorized emergency medical response training program. The devices are indicated for use on patients greater than 8 years old or over 55 lbs (25 kg) when used with the adult Pad-Pak™ (Pad-Pak-01 or Pad-Pak-07). They are indicated for use on children between 1 and 8 years of age or up to 55 lbs (25 kg) when used with the Pediatric-Pak™ (Pad-Pak-02).

### CONTRAINDICATION


If the patient is responsive or conscious, do not use the HeartSine samaritan PAD to provide treatment.

### WARNINGS

#### AEDs:

- The HeartSine samaritan PAD delivers therapeutic electrical shocks that can cause serious harm to either users or bystanders. Take care to ensure that no one touches the patient when a shock is to be delivered.
- Touching the patient during the analysis phase of treatment can cause interference with the diagnostic process. Avoid contact with the patient while the HeartSine samaritan PAD is analyzing the patient. The device will instruct you when it is safe to touch the patient.
- Do not delay treatment trying to find out the patient's exact age and weight. If a Pediatric-Pak or an alternative suitable defibrillator is not available, you may use an adult Pad-Pak.
- The SAM 360P is a fully automatic defibrillator. When required, it will deliver a shock to the patient WITHOUT user intervention.
- The SAM 450P CPR Rate Advisor is currently only intended to provide feedback on adult patients. If you treat a pediatric patient with the SAM 450P and an adult Pad-Pak, ignore any voice prompts regarding the rate of CPR.
- Do NOT use the HeartSine samaritan PAD in the vicinity of explosive gases, including flammable anesthetics or concentrated oxygen.
- Do NOT open or repair the device under any circumstances as there could be danger of electric shock. If damage is suspected, immediately replace the HeartSine samaritan PAD.

#### Pad-Paks:

- Do not use if the gel is dry.
- The Pediatric Pad-Pak is not for use on patients under 1 year old. For use with children up to the age of 8 years or up to 55 lbs (25 kg). DO NOT DELAY THERAPY IF YOU ARE NOT SURE OF EXACT AGE OR WEIGHT.
- Only HeartSine samaritan PADs with the  label are suitable for use with the Pediatric-Pak. If the HeartSine samaritan PAD you are using does not have this label, use the adult Pad-Pak if no alternatives are available.
- The use of the Pediatric-Pak will enable delivery of 50J shocks to the pediatric patient.
- The Pediatric-Pak contains a magnetic component (surface strength 6500 gauss). Avoid storage next to magnetically sensitive storage media. It is advised that Pediatric-Paks are stored separately when not in use.
- Never charge, short circuit, puncture, deform, incinerate, heat above 85°C or expose contents of TSO (Aviation) Pad-Pak to water. Remove when discharged.

## PRECAUTIONS

### AEDs:

- Proper placement of the HeartSine samaritan PAD electrode pads is critical. Electrode pads must be at least 1 in (2.5 cm) apart and should never touch one another.
- Do not use electrode pads if pouch is not sealed.
- Check the device periodically in accordance with the service and maintenance instructions provided in the User Manual.
- Operate the HeartSine samaritan PAD at least 6 feet (2 meters) away from all radio frequency devices or switch off any equipment causing interference.
- Use of the device outside the operating and storage ranges specified in the User Manual may cause the device to malfunction or reduce the shelf life of the Pad-Pak.
- Do not immerse any part of the HeartSine samaritan PAD in water or any type of fluid.
- Do not turn on the device unnecessarily as this may reduce the standby life of the device.
- Do not use any unauthorized accessories with the device as the HeartSine samaritan PAD may malfunction if non-approved accessories are used.
- Dispose of the device in accordance with national or local regulations.
- Check with the relevant local government health department for information about any requirements associated with ownership and use of a defibrillator in the region where it is to be used.

### Pad-Paks:

- Check expiration date.

### Saver EVO™ Software:

- Download the complete HeartSine samaritan PAD memory prior to erasing it. This information should be stored safely for future reference. Ensure that only the events you want to delete have been selected prior to deleting. Once deleted from your computer's memory, events cannot be regenerated and all information will be lost.

### POTENTIAL ADVERSE EFFECTS

The potential adverse effects (e.g., complications) associated with the use of an automated external defibrillator include, but are not limited to, the following:

- Failure to identify shockable arrhythmia.
- Failure to deliver a defibrillation shock in the presence of VF or pulseless VT, which may result in death or permanent injury.
- Inappropriate energy which could cause failed defibrillation or post-shock dysfunction.
- Myocardial damage.
- Fire hazard in the presence of high oxygen concentration or flammable anesthetic agents.
- Incorrectly shocking a pulse-sustaining rhythm and inducing VF or cardiac arrest.
- Bystander shock from patient contact during defibrillation shock.
- Interaction with pacemakers.
- Skin burns around the electrode placement area.
- Allergic dermatitis due to sensitivity to materials used in electrode construction.
- Minor skin rash.

### CAUTION

U.S. Federal law restricts this device to sale by or on the order of a physician.

Please consult the [User Manual at www.heartsine.com](#) for the complete list of indications, contraindications, warnings, precautions, potential adverse events, safety and effectiveness data, instructions for use and other important information.

1. Valenzuela TD, et al. 2000. Outcomes of Rapid Defibrillation by Security Officers After Cardiac Arrest in Casinos. *New England Journal of Medicine*. 343:1206-09.

## EMEA/APAC

HeartSine Technologies, Ltd.  
203 Airport Road West  
Belfast, Northern Ireland  
BT3 9ED  
Tel: +44 28 9093 9400  
Fax: +44 28 9093 9401  
info@heartsine.com



## U.S./Americas

HeartSine Technologies LLC  
121 Friends Lane, Suite 400  
Newtown, PA 18940  
Toll Free: (866) 478 7463  
Tel: +1 215 860 8100  
Fax: +1 215 860 8192  
info@heartsine.com



The HeartSine AEDs described in this brochure meet the European Medical Directive requirement.

UL Classified.  
See complete marking on HeartSine AED.

H009-032-338-3

**CAUTION:** U.S. Federal law restricts this device to sale by or on the order of a licensed practitioner.

© 2017 HeartSine Technologies LLC. All rights reserved.

[www.heartsine.com](http://www.heartsine.com)

